



We can provide interpreters for a variety of languages, information in larger print or other formats (e.g. audio) - please call us on 01932 723553.

You can use Text Relay to turn telephone communications into text.
Use the Relay UK app on your phone, tablet or PC.
You can also use Relay UK via Minicom or Uniphone.

اگر نیاز به ترجمہ دارید، لطفاً با شماره 01932 723553 تماس بگیرید۔

ਜੇ ਤੁਹਾਨੂੰ ਤਰਜਮੇ ਦੀ ਲੋੜ ਹੈ ਤਾਂ ਕਿਰਪਾ ਕਰਕੇ ਇਸ ਨੰਬਰ 'ਤੇ ਫੋਨ ਕਰੋ: 01932 723553

اگر آپ اس کا اردو زبان میں ترجمہ چاہتے ہیں، تو براہ کرم اس فون نمبر 01932 723553 پر رابطہ کریں

Se precisa de uma tradução por favor contacte: 01932 723553

আপনার অনুবাদের দরকার হলে এখানে যোগাযোগ করুন : 01932 723553

यदि आपको अनुवाद की ज़रूरत है तो कृपया इस नंबर पर फोन करें: 01932 723553

Jeżeli chcemy, aby te informacje w innym języku, proszę zadzwonić 01932 723553

Ashford Hospital
London Road
Ashford, Middlesex
TW15 3AA
Tel: **01784 884488**

St. Peter's Hospital
Guildford Road
Chertsey, Surrey
KT16 0PZ.
Tel: **01932 872000**

Website: www.ashfordstpeters.nhs.uk

Ambulatory Blood Pressure Monitoring

Clinical Measurements Department



What is an Ambulatory Blood Pressure monitor?

This is a portable Blood Pressure (BP) monitoring system which can monitor your BP whilst you carry on with your daily activities. It automatically records your BP over a period of time - usually for 24 hours.

Why is the test done?

The test is often requested by the doctor if they suspect high or low BP. This can be associated with symptoms such as dizziness and other symptoms which may come and go.

What does the test involve?

The test takes about 15-20 minutes. ***Please ask a member of staff if you require a chaperone during your appointment.***

A Cardiographer may ask you to remove your clothing from your top half in order for the monitor to be fitted comfortably.

The monitor is approximately the size of an old Walkman cassette player.

A BP cuff will be fitted around your arm which is attached to the monitor with a rubber tube. This fits into a pouch which is fitted to your own belt if you are wearing one or we can provide one. The monitor will be started by an initial BP reading. The monitor will not display any results throughout the 24-hour period. The cuff will inflate every 30 minutes during the day and from 22:00 / 10:00pm the cuff will inflate every hour until the following morning where it will revert back to every 30 minutes. When you feel the cuff start to inflate, please try to relax the arm to allow the monitor to get an accurate reading.

How do I prepare for the test?

Please wear loose-fitting clothes that are easy to put on and take off.

What do I have to do?

You will need to wear the monitor continuously for the required duration (including whilst sleeping at night). Carry out your usual daily activities whilst wearing the monitor. However, no bath or shower is permitted as the monitor **MUST NOT** get wet.

What happens after the test?

You will be provided with, and asked to sign, a waiver to return the monitor. If the monitor is lost or damaged, you may be charged for a replacement. You will need to return the monitor to the **ECG Department at Ashford Hospital** on the day that you have been instructed to do so. Please call the department on **01784 884266** if you have any concerns.

The results will be sent to the Consultant who requested the test. The results will be reviewed, and a letter will be sent to your General Practitioner (GP) advising them of the results. You will receive a copy of this letter. This may take 2-4 weeks. If you have not received your letter within 6 weeks, please contact your GP.

Further Information

We endeavor to provide an excellent service at all times, but should you have any concerns please, in the first instance, raise these with the Manager on duty. If they cannot resolve your concern, please contact our Patient Advice and Liaison Service (PALS) on 01932 723553 or email asp-tr.patient.advice@nhs.net. If you remain concerned, PALS can also advise upon how to make a formal complaint.

Author: C Camargo / K Garrett

Department: Clinical Measurements

Version: 3

Published: April 2024

Review: April 2026